

문서 번호	문서 제목	Version
AJ_DOC-116	Certification of Analysis for Sample Dilution Buffer/Na	1.1

제품명: Sample Dilution Buffer/Na




1.Product Information

제품명	: Sample Dilution Buffer/Na
Cat No.	: 60 01 007
Lot. No.	: 2605001
Source Batch.1	: Sodium Citrate tribasic dihydrate, BCCM0538
Source Batch.2	: Citric Acid, 0000351281
Water Batch	: AJ_HyperpureX_260511
CAS No.	: [Tri-Sodium Citrate dihydrate 6132-04-3], [Citric Acid Monohydrate 5949-29-1] [Hydrochloric Acid, 32 % 7647-01-0], [Octanoic Acid 124-07-2]
제조일자	: 2026년 05월 11일
보관 기한	: ~ 2028년 05월 11일
사용 기한	: 개봉 후 4주 이내
보관 조건	: 직사광선을 피하고 밀봉된 상태로 상온 보관(<25 °C)

3.Test Criteria

평가 항목	조건	결과
색상	: 투명한 액체	PASS
pH	: 2.20 ± 0.02	PASS
Retention Time of [ASP]	: 8.30 ~ 10.80	PASS
Ammonia Concentration	: < 1 ppm	PASS

3. 검수인

직함	제조사	평가일	최종 승인
이름	고승현	고승현	하성웅
날짜	260512	260512	260512
서명			

본 제품은 실험실 내 실험의 목적으로만 사용되어야 합니다. 가정용 혹은 약물 등으로 사용할 수 없습니다.
반드시 적합한 목적에 사용되어야 합니다.

(주)아주과학은 본 제품의 생산 및 평가 기간 동안 적합한 절차를 통해 시험이 진행되었으며 그에 따라 제품에 대한 품질을 보증합니다. 모든 제품은 적합한 절차와 목적에 맞게 사용되어야 하며, 이 외의 사용에 대해서는 (주)아주과학은 그 어떠한 책임도 지지 않습니다.